



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

H.A

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/517,307	08/22/2005	Karl-Heinz Banholzer	BANH3004/FJD	1943

23364 7590 04/03/2007  
BACON & THOMAS, PLLC  
625 SLATERS LANE  
FOURTH FLOOR  
ALEXANDRIA, VA 22314

EXAMINER
----------

ALLEN, ANDRE J

ART UNIT	PAPER NUMBER
----------	--------------

2855

SHORTENED STATUTORY PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE
3 MONTHS	04/03/2007	PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

**Office Action Summary**

Application No.

10/517,307

Applicant(s)

BANHOLZER ET AL.

Examiner

Andre J. Allen

Art Unit

2855

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 22 December 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 6-10 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 6-10 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some \* c) ☐ None of:
1. ☒ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- |  |   |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)   | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                       | 5) <input type="checkbox"/> Notice of Informal Patent Application                       |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)<br>Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____  |

## DETAILED ACTION

### ***Claim Rejections - 35 USC § 102***

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 6-10 are rejected under 35 U.S.C. 102(b) as being anticipated by Frick et al (US 6295875).

Regarding claim 6 Frick et al teaches a metal housing 114, and a ceramic pressure sensor (col. 2 lines 20-25) held in said metal housing, wherein: the surfaces of said metal housing which come in contact with a medium (abstract), whose pressure is to be measured, during measurement, are provided with a coating of one enamel and a glassy material (col. 6 lines 60-67).

Regarding claim 7 Frick et al teaches said housing is a flange (fig. 5) to be secured at a measuring location (fig. 5), said ceramic pressure sensor is

installed in said flange, and surfaces of said flange which come in contact with the medium are provided with said coating (col. 6 lines 60-67).

Regarding claim 8 Frick et al teaches said housing has a process connection 38, and surfaces of said process connection 38 which come in contact with the medium at a measurement location are provided with said coating (col. 6 lines 60-67).

Regarding claim 9 Frick et al teaches the pressure sensor is a differential pressure measuring cell 54; said housing has two side flanges (fig. 5), between which said ceramic pressure sensor is clamped, and surfaces of said side flanges which come in contact with the medium at the measurement location are provided with said coating (col. 6 lines 60-67).

Regarding claim 10 Frick teaches said housing 114 comprises steel or stainless steel (col. 6 line 64).

### ***Response to Arguments***

2. Applicant's arguments filed 12-22-06 have been fully considered but they are not persuasive.

In response to the applicants argument that the cited prior art does not disclose nor suggest surfaces of a metal housing coated with a coating of enamel or glassy material is not persuasive. That is, referring to the Frick et al patent, there is an explicit teaching of a housing 114 that includes cup like half wherein the halves include an insulating material (glass or enamel) that is fused to the metal housing (col. 6 lines 60-67). Therefore the arguments filed 12-22-06 are considered to be moot.

In response to the applicant's arguments that there is distinction between the process medium as claimed by the applicant and fluid within the Frick et al reference. Although the fluids/mediums used in both devices maybe different, the structural features in both the Frick et al reference and the applicants claim (a coated housing with enamel or glass) are both one in the same. Therefore a person in the filed of pressure sensor art would accomplish the same goal by insulating/protecting the housing from whatever hazardous environment that maybe encountered whether it be a fluid or medium. Furthermore, the fact that applicant has recognized another advantage which would flow naturally from following the suggestion of the prior art cannot be the basis for patentability when the differences would otherwise be obvious. See *Ex parte Obiaya*, 227 USPQ 58, 60 (Bd. Pat. App. & Inter. 1985).

***Conclusion***

**THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Andre J. Allen whose telephone number is 571-272-2174. The examiner can normally be reached on mon-fri 8:00-4:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Edward Lefkowitz can be reached on 571-272-

Art Unit: 2855

2180. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



André J. Allen  
Patent Examiner  
Art Unit 2855